



DERMANET® GTL SILICONE WOUND CONTACT LAYER

	NOT MADE WITH NATURAL RUBBER LATEX
	SINGLE PATIENT USE
	STERILE
RX ONLY	Federal U.S.A. law restricts this device to sale or use by or on the order of a physician or properly licensed practitioner.



IMPORTANT INFORMATION

Read carefully and completely before using or applying.

INTENDED USE

DeRoyal® Dermanet® GTL Silicone Wound Contact Layer can be used on infected and non-infected wounds, including dermal ulcers (e.g. leg ulcers, pressure ulcers), diabetic ulcers, abdominal wounds, superficial wounds, lacerations, cuts, surgical incision, abrasions, donor sites, tunneling wounds, deep wounds, and burn wounds.

INDICATIONS FOR USE

- Dermanet® wound dressings are indicated for minor abrasions, minor lacerations, and minor scalds and burns.
- When under the care or direction of a licensed practitioner they are indicated for:
 - Dermal ulcers (e.g., pressure ulcers)
 - Deep wounds
 - Tunneling wounds
 - Pressure ulcers
 - Venous ulcers
 - Dermal lesions (or secreting skin injuries)
 - Burn injuries
 - Donor sites
 - Lacerations
 - Surgical incisions

CONTRAINDICATIONS

Dermanet should not be used on:

- Third degree burns
- Individuals who are sensitive to or have had an allergic reaction to the dressing or its materials

WARNINGS

- Read and understand all warnings, cautions, and directions completely and carefully before use. Failure to do so may result in serious injury.

- Do not reuse, reprocess, or re-sterilize this dressing. This device is for single use only. Reuse, reprocessing, or re-sterilization may damage the dressing, and may result in injury, infection, or death.
- Do not use if the individual package is opened or if the package or the dressing is damaged. Carefully inspect the dressing for damage or defects prior to use.
- Seek medical attention if your wound is more severe than a minor abrasion, minor laceration, or minor scald and burn. Also seek medical attention if you have signs of an infection or allergic reaction such as fever, redness, pain, or swelling.
- Do NOT use dressing for surgical implantation
- Ensure dressing is not left behind in wound bed post wound assessment
- Before application, ensure dressing does not contain foreign particulate or debris.
- Do not use if product is ripped, torn, or damaged.
- Constantly monitor wound and do not leave a dressing on for an extended period of time. Refer to a licensed practitioner for frequency of dressing change.

CAUTIONS

- If you have any signs of an allergic reaction, discontinue using the dressing and rinse the wound and surrounding skin.
- Ensure the dressing fully covers your wound.
- Prior to removal, ensure the dressing has not adhered to the wound. If the dressing has adhered to the wound, rinse with water or saline solution to loosen.
- If dressing comes off wound, discard and replace.
- Ensure dressing allows exudate to flow through.
- Do not apply dressing with too much pressure on wounds with potential for high granulating tissue or meshed skin grafts.



DERMANET® GTL SILICONE WOUND CONTACT LAYER

	NOT MADE WITH NATURAL RUBBER LATEX
	SINGLE PATIENT USE
	STERILE
RX ONLY	Federal U.S.A. law restricts this device to sale or use by or on the order of a physician or properly licensed practitioner.



IMPORTANT INFORMATION

Read carefully and completely before using or applying.

INTENDED USE

DeRoyal® Dermanet® GTL Silicone Wound Contact Layer can be used on infected and non-infected wounds, including dermal ulcers (e.g. leg ulcers, pressure ulcers), diabetic ulcers, abdominal wounds, superficial wounds, lacerations, cuts, surgical incision, abrasions, donor sites, tunneling wounds, deep wounds, and burn wounds.

INDICATIONS FOR USE

- Dermanet® wound dressings are indicated for minor abrasions, minor lacerations, and minor scalds and burns.
- When under the care or direction of a licensed practitioner they are indicated for:
 - Dermal ulcers (e.g., pressure ulcers)
 - Deep wounds
 - Tunneling wounds
 - Pressure ulcers
 - Venous ulcers
 - Dermal lesions (or secreting skin injuries)
 - Burn injuries
 - Donor sites
 - Lacerations
 - Surgical incisions

CONTRAINDICATIONS

Dermanet should not be used on:

- Third degree burns
- Individuals who are sensitive to or have had an allergic reaction to the dressing or its materials


WARNINGS

- Read and understand all warnings, cautions, and directions completely and carefully before use. Failure to do so may result in serious injury.

- Do not reuse, reprocess, or re-sterilize this dressing. This device is for single use only. Reuse, reprocessing, or re-sterilization may damage the dressing, and may result in injury, infection, or death.
- Do not use if the individual package is opened or if the package or the dressing is damaged. Carefully inspect the dressing for damage or defects prior to use.
- Seek medical attention if your wound is more severe than a minor abrasion, minor laceration, or minor scald and burn. Also seek medical attention if you have signs of an infection or allergic reaction such as fever, redness, pain, or swelling.
- Do NOT use dressing for surgical implantation
- Ensure dressing is not left behind in wound bed post wound assessment
- Before application, ensure dressing does not contain foreign particulate or debris.
- Do not use if product is ripped, torn, or damaged.
- Constantly monitor wound and do not leave a dressing on for an extended period of time. Refer to a licensed practitioner for frequency of dressing change.

CAUTIONS

- If you have any signs of an allergic reaction, discontinue using the dressing and rinse the wound and surrounding skin.
- Ensure the dressing fully covers your wound.
- Prior to removal, ensure the dressing has not adhered to the wound. If the dressing has adhered to the wound, rinse with water or saline solution to loosen.
- If dressing comes off wound, discard and replace.
- Ensure dressing allows exudate to flow through.
- Do not apply dressing with too much pressure on wounds with potential for high granulating tissue or meshed skin grafts.



DIRECTIONS FOR USE

APPLICATION OF DERMANET

1. Thoroughly cleanse wound.
2. Remove Dermanet from packaging
3. Carefully apply Dermanet to wound
4. If applicable, topical medications can be applied directly to Dermanet and then placed over the wound
5. Cover with appropriate secondary dressing (e.g. gauze, Covaderm[®], Polyderm[™], etc...) based upon amount of wound drainage
6. Secure dressing in place as needed with tape, roll gauze or Stretch Net[™].

REMOVAL OF DERMANET

1. To remove Dermanet, remove secondary cover dressing
2. Irrigate wound with sterile saline to help loosen dressing if required
3. Reapply Dermanet as necessary
4. Dispose of soiled dressing per institution's policy to avoid exposure to or contamination of biohazard.

STORAGE AND TRANSPORT CONDITIONS

Store at room temperature and avoid excessive heat and humidity.

	KEEP AWAY FROM SUNLIGHT
	KEEP DRY

WARRANTY

DeRoyal products are warranted for ninety (90) days from the date of shipment from DeRoyal as to product quality and workmanship. **DEROYAL'S WRITTEN WARRANTIES ARE GIVEN IN LIEU OF ANY IMPLIED WARRANTIES, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.**



Manufacturer


DeRoyal Industries, Inc.
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#757347

Revised 2/17

©2017 DeRoyal Industries, Inc.

All Rights Reserved. DeRoyal, the DeRoyal logo, Dermanet, Covaderm, Polyderm, StretchNet and Improving Care. Improving Business. are trademarks or registered trademarks of DeRoyal Industries, Inc.



DIRECTIONS FOR USE

APPLICATION OF DERMANET

1. Thoroughly cleanse wound.
2. Remove Dermanet from packaging
3. Carefully apply Dermanet to wound
4. If applicable, topical medications can be applied directly to Dermanet and then placed over the wound
5. Cover with appropriate secondary dressing (e.g. gauze, Covaderm[®], Polyderm[™], etc...) based upon amount of wound drainage
6. Secure dressing in place as needed with tape, roll gauze or Stretch Net[™].

REMOVAL OF DERMANET

1. To remove Dermanet, remove secondary cover dressing
2. Irrigate wound with sterile saline to help loosen dressing if required
3. Reapply Dermanet as necessary
4. Dispose of soiled dressing per institution's policy to avoid exposure to or contamination of biohazard.

STORAGE AND TRANSPORT CONDITIONS

Store at room temperature and avoid excessive heat and humidity.

	KEEP AWAY FROM SUNLIGHT
	KEEP DRY

WARRANTY

DeRoyal products are warranted for ninety (90) days from the date of shipment from DeRoyal as to product quality and workmanship. **DEROYAL'S WRITTEN WARRANTIES ARE GIVEN IN LIEU OF ANY IMPLIED WARRANTIES, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.**



Manufacturer

DeRoyal Industries, Inc.
200 DeBusk Lane
Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

Part#757347

Revised 2/17

©2017 DeRoyal Industries, Inc.

All Rights Reserved. DeRoyal, the DeRoyal logo, Dermanet, Covaderm, Polyderm, StretchNet and Improving Care. Improving Business. are trademarks or registered trademarks of DeRoyal Industries, Inc.