

Tracker EX Walker

Single Patient Use Only

Rx Only

IMPORTANT INFORMATION

Please read all instructions before use. Correct application is essential to proper functioning of product.

INDICATIONS


Grade 2 and 3 ankle sprains; stable fractures of tibia, fibula, talus and tarsus; post operative immobilization (soft tissue reconstructions and open reduction internal fixation); metatarsal fractures

PRECAUTIONS

- This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing use instructions and precautions to other healthcare practitioners, care providers involved in the patients' care and the patient.
- Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

INSTRUCTIONS FOR USE

1. Remove inner liner/ wrap from brace. Position foot and lower leg into the liner.
2. Wrap snugly around limb. Fasten with "hook" fasteners to the outside of the liner. Insert foot and lower leg into walker.
3. Carefully center the uprights (side panels) over the ankle to ensure a comfortable fit and proper support. Remove plastic covers from uprights and secure hook to liner.
4. Fasten walker using hook and loop leg straps. Begin fastening at the bottom and work your way up the leg. **NOTE:** Readjust as needed for a secure and comfortable fit.



Tracker EX Walker

Single Patient Use Only

Rx Only

IMPORTANT INFORMATION

Please read all instructions before use. Correct application is essential to proper functioning of product.

INDICATIONS


Grade 2 and 3 ankle sprains; stable fractures of tibia, fibula, talus and tarsus; post operative immobilization (soft tissue reconstructions and open reduction internal fixation); metatarsal fractures

PRECAUTIONS

- This product is to be fitted initially by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing use instructions and precautions to other healthcare practitioners, care providers involved in the patients' care and the patient.
- Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

INSTRUCTIONS FOR USE

1. Remove inner liner/ wrap from brace. Position foot and lower leg into the liner.
2. Wrap snugly around limb. Fasten with "hook" fasteners to the outside of the liner. Insert foot and lower leg into walker.
3. Carefully center the uprights (side panels) over the ankle to ensure a comfortable fit and proper support. Remove plastic covers from uprights and secure hook to liner.
4. Fasten walker using hook and loop leg straps. Begin fastening at the bottom and work your way up the leg. **NOTE:** Readjust as needed for a secure and comfortable fit.





WASHING INSTRUCTIONS

Liner should be hand washed using cold water and mild detergent. Air dry.



Distributor
DeRoyal
200 DeBusk Lane, Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

©2006 DeRoyal. All Rights Reserved.
Revised 12/06



WASHING INSTRUCTIONS

Liner should be hand washed using cold water and mild detergent. Air dry.



Distributor
DeRoyal
200 DeBusk Lane, Powell, TN 37849 U.S.A.
888.938.7828
865.938.7828

©2006 DeRoyal. All Rights Reserved.
Revised 12/06