

EN: ANKLE CONTRACTURE BOOT


MANUFACTURED BY:
DeRoyal Industries, Inc.
200 DeBusk Lane
Powell, TN 37849 USA
888.938.7828 or (001) 865.938.7828
www.deroyal.com

MedEnvoy Global B.V.
Prinses Margrietplantsoen 33 – Suite 123
2595 AM The Hague
The Netherlands

UK Responsible Person
MedEnvoy UK Limited
85, Great Portland Street, First Floor
London, W1W 7LT
United Kingdom



PART#1-4300G I REVISED 10/23

©2023 DeRoyal Industries, Inc.
All Rights Reserved. DeRoyal and the DeRoyal logo
are registered trademarks of DeRoyal Industries, Inc.

ES: BOTA PARA CONTRACTURAS DE TOBILLO

FR: BOTTE POUR CONTRACTURES DE CHEVILLE

DE: STIEFEL FÜR KONTRAKTUREN DES SPRUNGGelenKS

IT: STIVALETTO PER CONTRATTURA CAVIGLIA

PT: BOTTE DE CONTRATURA DO TORNozelo

NL: LAARS VOOR ENKELCONTRACTUUR

SE: ORTOS FÖR FOTLEDSKONTRAKTUR

TR: AYAK BİLEĞİ KONTRAKTÜRKÜ BOTU

EL: ΜΠΟΤΑ ΑΝΤΙΜΕΤΩΠΙΣΗΣ ΣΥΣΠΑΣΗΣ ΠΟΔΟΚΝΗΜΙΚΗΣ

ANKLE CONTRACTURE BOOT

	SINGLE PATIENT USE
	NON-Sterile
	MEDICAL DEVICE
	NOT MADE WITH NATURAL RUBBER LATEX
RX ONLY	FEDERAL U.S.A. LAW RESTRICTS THIS DEVICE TO SALE OR USE BY OR ON THE ORDER OF A PHYSICIAN OR PROPERLY LICENSED PRACTITIONER.

IMPORTANT INFORMATION

Please read all instructions, warnings and cautions before use. Correct application is essential for proper product function and to reduce the risk of injury or re-injury inherent with the use of any brace. Use only on the person it was provided to by a healthcare professional and only for the use it was intended.

INTENDED USE
The DeRoyal® Ankle Contracture Boot is intended to be used to help prevent foot rotation, foot drop and heel decubitus ulcers by restricting lateral foot rotation, and off-loading the heel, the Achilles, and the calf.

CONTRAINdicATIONS
Do not use on patients with severe plantar flexion contractures, severe spasticity or thrombosis.

WARNINGS
• Inspect device for damaged or missing components prior to patient application.
• Consult physician or properly licensed practitioner immediately if you experience sensation changes, unusual reactions, swelling or increased pain while using this product.
• If device is used with a sequential/intermittent compression device, be sure tubing does not interfere with off-loading of the patient's foot, ankle or calf.
• Discontinue use of the device and consult physician or properly licensed practitioner if it becomes damaged in any way, limiting its ability to perform as intended.

CAUTIONS
A physician or properly licensed practitioner (a "prescriber") who is familiar with the use and purpose of this brace must fit it to the user. The prescriber has a duty to provide written instructions and risks related to the use of this brace to other healthcare practitioners treating the users and the users themselves, including duration of use. The instructions provided in this sheet do not supersede hospital protocol or direct orders of the prescriber.

• Inspect that fit of the brace on the patient to ensure fit is proper.

OVERTIGHTENING
• Consult your prescriber immediately if you experience sensitivity changes, unusual reactions, swelling or increased pain while using this brace. Discontinue use if pressure injuries develop.

• Take special care if the user is diabetic or has poor circulation. These users may have decreased skin sensitivity and are at greater risk for poor peripheral circulation and pressure injuries.

• Ankle Contracture boot is recommended for light gait training. Patient can only transfer less than 10 feet. Ensure sole is

attached prior to ambulation.
• Be sure to follow your facility's policies and guidelines for frequency of patient monitoring.

INSTRUCTIONS FOR USE

- Measure the circumference of the calf 6 inches up from the bottom of the heel and around the widest part of the foot to determine the appropriate size. Please refer to the size chart below.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE TOE PIECE
Adjust the length of the toe piece by loosening the thumb knob on the underside of the sole. Position the toe piece to the desired length and retighten the knob. The toe piece can also be rotated to either side of the foot. **NOTE:** The thumb knob does not detach from the boot.

ATTACHING AND REMOVING THE BOOT SOLE
To apply the sole, place the boot sole on the bottom of the frame, aligning the hole in the sole with the thumb knob. With thumbs on either side of the hole, push the sole down and toward the heel. The thumb knob will slide to the back of the slot on the sole. Push down on the sole at the heel until it locks into place.

REPLACING LINERS
The single piece foam liner comfortably wraps around the foot and ankle and allows airflow. Hook and loop closures make donning and doffing of the boot easy. The single piece fleece liner offers maximum comfort for the patient. This is especially important for patients with a tendency of skin shear.

REPLACEMENT LINERS
1. Remove the liners, pull off the toe-piece extender and anti-rotation bar covers. Unfasten the hook and loop closure at the back of the heel. Remove the toe-piece extender from the boot

frame and pull the foot cuff forward off the plastic frame.

- To apply the liners, slide the calf of the frame into the calf pocket of the liner. Slide the foot off the frame into the foot pocket of the liner. Secure the hook and loop closure at the back of the heel. Slide the covers over the anti-rotation bar and toe piece.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

PLEASE NOTE: For X-Small (Size A) device, liner pieces cannot be removed. Spot clean the device with mild fabric detergent and water using a non-linting wipe or utilize a hospital approved disinfectant wipe. **DO NOT** submerge device in water. Allow device to dry completely prior to patient re-application.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE ANTI-ROTATION BAR
Rotate the bar clockwise or counter clock-wise to the desired position. This will assist in the prevention of inversion or eversion. To apply the high abduction bar, remove the cover from the anti-rotation bar, fully rotate the anti-rotation bars to the side, slide the abduction bar over both anti-rotation bars and lock into position. To release the high abduction bar, pull up on the lever and slide the abduction bar off the anti-rotation bar.

ATTACHING AND REMOVING THE BOOT SOLE
To apply the sole, place the boot sole on the bottom of the frame, aligning the hole in the sole with the thumb knob. With thumbs on either side of the hole, push the sole down and toward the heel. The thumb knob will slide to the back of the slot on the sole. Push down on the sole at the heel until it locks into place.

REPLACING LINERS
The single piece foam liner comfortably wraps around the foot and ankle and allows airflow. Hook and loop closures make donning and doffing of the boot easy. The single piece fleece liner offers maximum comfort for the patient. This is especially important for patients with a tendency of skin shear.

REPLACEMENT LINERS
1. Remove the liners, pull off the toe-piece extender and anti-rotation bar covers. Unfasten the hook and loop closure at the back of the heel. Remove the toe-piece extender from the boot

frame and pull the foot cuff forward off the plastic frame.

- To apply the liners, slide the calf of the frame into the calf pocket of the liner. Slide the foot off the frame into the foot pocket of the liner. Secure the hook and loop closure at the back of the heel. Slide the covers over the anti-rotation bar and toe piece.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

PLEASE NOTE: For X-Small (Size A) device, liner pieces cannot be removed. Spot clean the device with mild fabric detergent and water using a non-linting wipe or utilize a hospital approved disinfectant wipe. **DO NOT** submerge device in water. Allow device to dry completely prior to patient re-application.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE ANTI-ROTATION BAR
Rotate the bar clockwise or counter clock-wise to the desired position. This will assist in the prevention of inversion or eversion. To apply the high abduction bar, remove the cover from the anti-rotation bar, fully rotate the anti-rotation bars to the side, slide the abduction bar over both anti-rotation bars and lock into position. To release the high abduction bar, pull up on the lever and slide the abduction bar off the anti-rotation bar.

ATTACHING AND REMOVING THE BOOT SOLE
To apply the sole, place the boot sole on the bottom of the frame, aligning the hole in the sole with the thumb knob. With thumbs on either side of the hole, push the sole down and toward the heel. The thumb knob will slide to the back of the slot on the sole. Push down on the sole at the heel until it locks into place.

REPLACING LINERS
The single piece foam liner comfortably wraps around the foot and ankle and allows airflow. Hook and loop closures make donning and doffing of the boot easy. The single piece fleece liner offers maximum comfort for the patient. This is especially important for patients with a tendency of skin shear.

REPLACEMENT LINERS
1. Remove the liners, pull off the toe-piece extender and anti-rotation bar covers. Unfasten the hook and loop closure at the back of the heel. Remove the toe-piece extender from the boot

frame and pull the foot cuff forward off the plastic frame.

- To apply the liners, slide the calf of the frame into the calf pocket of the liner. Slide the foot off the frame into the foot pocket of the liner. Secure the hook and loop closure at the back of the heel. Slide the covers over the anti-rotation bar and toe piece.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

PLEASE NOTE: For X-Small (Size A) device, liner pieces cannot be removed. Spot clean the device with mild fabric detergent and water using a non-linting wipe or utilize a hospital approved disinfectant wipe. **DO NOT** submerge device in water. Allow device to dry completely prior to patient re-application.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE ANTI-ROTATION BAR
Rotate the bar clockwise or counter clock-wise to the desired position. This will assist in the prevention of inversion or eversion. To apply the high abduction bar, remove the cover from the anti-rotation bar, fully rotate the anti-rotation bars to the side, slide the abduction bar over both anti-rotation bars and lock into position. To release the high abduction bar, pull up on the lever and slide the abduction bar off the anti-rotation bar.

ATTACHING AND REMOVING THE BOOT SOLE
To apply the sole, place the boot sole on the bottom of the frame, aligning the hole in the sole with the thumb knob. With thumbs on either side of the hole, push the sole down and toward the heel. The thumb knob will slide to the back of the slot on the sole. Push down on the sole at the heel until it locks into place.

REPLACING LINERS
The single piece foam liner comfortably wraps around the foot and ankle and allows airflow. Hook and loop closures make donning and doffing of the boot easy. The single piece fleece liner offers maximum comfort for the patient. This is especially important for patients with a tendency of skin shear.

REPLACEMENT LINERS
1. Remove the liners, pull off the toe-piece extender and anti-rotation bar covers. Unfasten the hook and loop closure at the back of the heel. Remove the toe-piece extender from the boot

frame and pull the foot cuff forward off the plastic frame.

- To apply the liners, slide the calf of the frame into the calf pocket of the liner. Slide the foot off the frame into the foot pocket of the liner. Secure the hook and loop closure at the back of the heel. Slide the covers over the anti-rotation bar and toe piece.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

PLEASE NOTE: For X-Small (Size A) device, liner pieces cannot be removed. Spot clean the device with mild fabric detergent and water using a non-linting wipe or utilize a hospital approved disinfectant wipe. **DO NOT** submerge device in water. Allow device to dry completely prior to patient re-application.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE ANTI-ROTATION BAR
Rotate the bar clockwise or counter clock-wise to the desired position. This will assist in the prevention of inversion or eversion. To apply the high abduction bar, remove the cover from the anti-rotation bar, fully rotate the anti-rotation bars to the side, slide the abduction bar over both anti-rotation bars and lock into position. To release the high abduction bar, pull up on the lever and slide the abduction bar off the anti-rotation bar.

ATTACHING AND REMOVING THE BOOT SOLE
To apply the sole, place the boot sole on the bottom of the frame, aligning the hole in the sole with the thumb knob. With thumbs on either side of the hole, push the sole down and toward the heel. The thumb knob will slide to the back of the slot on the sole. Push down on the sole at the heel until it locks into place.

REPLACING LINERS
The single piece foam liner comfortably wraps around the foot and ankle and allows airflow. Hook and loop closures make donning and doffing of the boot easy. The single piece fleece liner offers maximum comfort for the patient. This is especially important for patients with a tendency of skin shear.

REPLACEMENT LINERS
1. Remove the liners, pull off the toe-piece extender and anti-rotation bar covers. Unfasten the hook and loop closure at the back of the heel. Remove the toe-piece extender from the boot

frame and pull the foot cuff forward off the plastic frame.

- To apply the liners, slide the calf of the frame into the calf pocket of the liner. Slide the foot off the frame into the foot pocket of the liner. Secure the hook and loop closure at the back of the heel. Slide the covers over the anti-rotation bar and toe piece.
- Open all hook and loop closures on the boot. Flex the knee to approximately 45°. Place the foot into the boot. Make sure the sole of the foot is resting on the bottom of the boot and the calf of the leg is resting on the calf portion of the boot. **NOTE:** The heel should not come in contact with the frame, but should be "floating" in the boot.
- Wrap the left side of the foot cuff over the instep. Wrap the right side of the foot cuff over the top of the left, and secure with the hook and loop.

PLEASE NOTE: For X-Small (Size A) device, liner pieces cannot be removed. Spot clean the device with mild fabric detergent and water using a non-linting wipe or utilize a hospital approved disinfectant wipe. **DO NOT** submerge device in water. Allow device to dry completely prior to patient re-application.

CLEANING AND/OR MAINTENANCE
Remove all line pieces and fasten all hook and loop closures. Machine wash liner with a mild fabric detergent in lukewarm water. Do not use bleach. Do not iron. Tumble dry on low. Boot frame and other hardware components can be wiped with a hospital approved disinfectant. Allow device to dry completely prior to patient re-application.

ADJUSTING THE ANTI-ROTATION BAR
Rotate the bar clockwise or counter clock-wise to the desired position. This will assist in the prevention of inversion or eversion. To apply the high abduction bar, remove the cover from the anti-rotation bar, fully rotate the anti-rotation bars to the side, slide the abduction bar over both anti-rotation bars and lock into position. To release the high abduction bar, pull up

STIVALETTO PER CONTRATTURA CAVIGLIA

	PER UTILIZZO SU UN UNICO PAZIENTE
	NON STERILE
	DISPOSITIVI MEDICI
	NON IN LATICE DI GOMMA NATURALE
	LE LEGGI FEDERALI DEGLI STATI UNITI D'AMERICA LIMITANO LA VENDITA DEL PRESENTE DISPOSITIVO A MEDICI, A PERSONALE AUTORIZZATO O A OPERATORI SANITARI ABILITATI.

INFORMAZIONI IMPORTANTI
Leggere tutte le istruzioni, le avvertenze e le precauzioni prima dell'uso. Applicare correttamente il prodotto è per garantire il regolare funzionamento e ridurre il rischio di lesioni o iniezioni nell'uso di qualsiasi tipo di foderi. Da utilizzare esclusivamente per la persona a cui è stato fornito dal professionista sanitario ed esclusivamente per l'uso a cui è destinato.

USO PREVISTO
Lo Stivaletto per contrattura caviglia DeRoyal® è concepito per contribuire a prevenire la rotazione del piede, la crisi del piede e la formazione di ulcere da decubito sul tallone limitando la rotazione laterale del piede e scardinando il tallone, il tendine d'Achille e il polpaccio.

CONTRAINDIÇAÇÕES
Non utilizzare su pazienti con gravi contratture in flessione plantare, grave spasticità o trombo.

AVVERTENZA
• Prima di applicare l'apparecchio al paziente, controllare che non presenti danni né componenti mancati.

• Rivolgervi immediatamente al medico curante o a operatori sanitari abilitati se si riscontrano cambiamenti, reazioni insolite, tumefazioni o dolore prolungato durante l'utilizzo di questo prodotto.

• Qualora forti venga utilizzata in combinazione con dispositivi di compressione sequenziale/intermittente, assicurarsi che il tubo non interferisca con lo scarico del piede, della caviglia o del polpaccio.

• Intervenire l'uso del dispositivo e consultare un medico o un operatore sanitario abilitato se il dispositivo dovesse subire danni tali da limitare la capacità di funzionare come previsto.

PRECAUZIONI
• Questo tutore deve essere applicato all'utilizzatore da un medico o da un operatore sanitario abilitato ("prescrittore") che ne conosca l'utilizzo e la finalità. Il prescrittore ha il dovere di fornire istruzioni per indossare il prodotto e informazioni sui rischi correlati all'uso di questo tutore ad altri operatori sanitari che hanno in cura gli utilizzatori e agli utilizzatori stessi, compresa la durata di utilizzo. Le istruzioni fornite in questa foglietto non sostituiscono il protocollo ospedaliero o le indicazioni cliniche del prescrittore.

• Verificare che il tutore si adatti perfettamente al paziente. **NON STRINGERE**.

• Rivolgervi immediatamente al prescrittore se si riscontrano cambiamenti, reazioni insolite, tumefazioni o dolore prolungato durante l'utilizzo di questo tutore. Sospendere l'uso qualora insorgano lesioni da pressione.

• Si raccomanda una particolare attenzione se l'utilizzatore è affetto da diabete o disturbi circolatori. Questi utilizzatori potrebbero avere una

ridotta sensibilità cutanea e sono quindi esposti a maggiori rischi di cattiva circolazione periferica e lesioni da pressione.

• Lo stivaletto per contrattura caviglia è consigliato per allungamento leggero di riduzione alla deambulazione. Il paziente è in grado di spostarsi per meno di 10 piedi (3 m). Assicurarsi che la suola sia fissata prima della deambulazione.

• Accertarsi di seguire i regolamenti della propria struttura sanitaria e le linee guida relative alla frequenza di monitoraggio del paziente.

ISTRUZIONI PER L'USO

1. Misurare la circonferenza del polpaccio 8 pollici (20,3 cm) sopra la parte inferiore del tallone e attorno alla parte più ampia del piede per determinare la dimensione appropriata. Fare riferimento alla tabella delle taglie riportata a pagina 4.

2. Aprire tutte le chiusure a strappo dello stivaletto. Flettere il ginocchio di circa 45°. Posizionare il piede nello stivaletto. Assicurarsi che la piattaforma dei piedi poggi sul fondo dello stivaletto e il polpaccio poggi contro l'apposita porzione dello stivaletto. **NOTA:** Il tallone non deve venire a contatto con il tallone "ma restare libero" nello stivaletto.

3. Avvolgere la parte sinistra della protezione del piede sul collo del polpaccio. Avvolgere la parte destra della protezione del piede con la parte superiore di quella sinistra, quindi fissare con la chiusura a strappo. **NOTA:** non stringere eccessivamente. Eseguire valutazioni periodiche delle condizioni della pelle e della circolazione del paziente secondo le procedure e i regolamenti ospedalieri.

REGOLAZIONE DEL SUPPORTO DITA

Regolare la lunghezza del supporto dita allentando la rotellina nella parte inferiore della suola. Posizionare il supporto dita in base alla lunghezza desiderata e serrare nuovamente la rotellina. Il supporto dita può essere inoltre ruotato su ciascun lato del piede. **NOTA:** La rotellina non può essere rimossa dallo stivaletto.

REGOLAZIONE DELLA BARRA ANTI-ROTATORE

Ruotare la barra in senso orario o in senso antiorario fino alla posizione desiderata. Tale barra contribuisce ad impedire movimenti di inversione o eversione. Per applicare la barra di abduzione dell'anca, rimuovere la copertura dalla barra anti-rotazione, ruotare completamente le barre anti-rotazione in posizione laterale, far scorrere la barra di abduzione al di sopra di entrambe le barre anti-rotazione e fissarla in posizione. Per rilasciare la barra di abduzione dell'anca, tirare la leva verso l'alto e far scorrere la barra di abduzione finché non si troverà più sulla barra anti-rotazione.

APPLICAZIONE E REMOZIONE DELLA SULLA DELLO STIVALETTO

Per applicare la suola dello stivaletto, posizionarla sul fondo del piede e allineare il foro presente nella suola alla rotellina. Tenendo i pollici su lati del foro, premere la suola verso il basso in direzione della talla. La rotellina scorrerà verso la parte posteriore della scalatura nella suola. Premerla nella corrispondenza del tallone fino a quando scatta in posizione.

Per rimuovere la suola, posizionare le dita fra la suola e il tallone in corrispondenza del tallone. Premere la base del palmo della mano sulla suola tirando via la suola dal tallone con le dita fino a sfibocarla. Far scorrere la suola verso le dita fino a quando la rotellina è fissa con il foro, quindi rimuovere la suola.

OPZIONE FODERA

La fodera in sughero monopezzo avvolge comodamente il piede e la caviglia consentendo la respirazione dell'aria. Le chiusure a strappo semplificano l'inserimento e la rimozione dello stivaletto. La fodera in vello monopezzo offre al paziente il massimo comfort. Ciò è particolarmente

tijdsens het gebruik van deze brace. Stop met het gebruik als er drukverwondingen ontstaan.

• Als de gebruiker diabetisch patiënt is of een slechte bloedcirculatie heeft, dien extra zorgvuldigheid in acht te worden genomen. Bij deze gebruikers kan de huid minder gevoelig zijn waardoor ze kans hebben op een slechte perifere doorbloeding en drukverwondingen.

• De laars voor enkelcontractuur wordt aanbevolen voor lichte loopenoefeningen. De patiënt kan er slechts drie meter mee lopen. Zorg dat de voet vastzit wanneer er met de laars wordt gelopen.

• Volg het beleid en de richtlijnen van uw instelling ten aanzien van de frequentie waarmee de patiënt moet worden gecontroleerd.

GARANZIA

I prodotti DeRoyal sono garantiti per cento giorni (120 giorni) da data d'acquisto dello stabilimento DeRoyal per quanto riguarda qualità e lavorazione del prodotto. LEGGIMI SCRITTE DI DERROYAL SOSTITUISCONO QUALSIASI GARANZIA IMPLICITA, IN COMPRESE LE GARANZIE DI COMMERCIALITÀ O IDONEITÀ PER UNO USO SPECIFICO.

vergankelijk het aan- en uitdoen van de laars. De enkelvoudige fleecedevoering biedt de patiënt maximale comfort. Dit is met name belangrijk voor patiënten die gevoelig zijn voor doorligwonden.

VOEGEN VERHANDELINGEN

1. Verwijder het teenstuk en de bescherming van de anti-rotatiestaaf om de voering te kunnen verwijderen. Maak de klettensluiting af aan de achterkant van de hiels. Verwijder het teenstuk van het laarsframe en trek de laars voorwaarts uit het plastic frame.

2. Trek het kuitgedeelte van de voering omhoog uit het frame.

3. Schuif het kuitgedeelte van het frame in de kuitopening van de voering om de voering te plaatsen. Schuif het voetgedeelte van het frame af in de voetopening van de voering. Bevestig de klettensluiting aan de achterkant van de hiel. Verwijder de beschermingen over de anti-rotatiestaaf en het teenstuk.

REINIGING EN VOEDING

Vervolgen alle voorzorgsmaatregelen voor de klettensluitingen. Wassen in de wasmachine in lauwwater met een mild wasmiddel voor textiel. Geen bleekmiddel gebruiken. Niet strijken. Op lange stand in de wasdroger drogen.

Het laarsframe en andere harde onderdelen kunnen worden afgewassen met een desinfectiemiddel dat door het ziekenhuis is geadviseerd.

4. Wikkel de linkerkant van de kuitbescherming over de kuit. Wikkel de rechterkant van de kuitbescherming over de bovenkant links, en bevestig deze met de klettensluiting.

5. Wikkel de linkerkant van de kuitbescherming over de kuit. Wikkel de rechterkant van de kuitbescherming over de bovenkant rechts, en bevestig deze met de klettensluiting.

6. Controleer na het wasen of het product nog intact is. Als het product is beschadigd of onvoldoende kan worden gereinigd, dient u het product niet opnieuw te gebruiken maar te vervangen door een nieuw product.

AFSTELLEN VAN HET TEENSTUK

Stel de lengte van het teenstuk in door de draaiknop aan de onderkant van de zoom los te maken. Plaats het teenstuk op de gevreesde lengte en draai de knop weer vast. Het teenstuk kan ook naar weerszijden van de voet worden gedraaid.

AFSTELLEN VAN DE ANTI-ROTATIESTAAF

Draai de staaf rechts- of linksonder om de gewenste stand. Dil het voorhoofd van de staaf. Voor het aanscherpen van de anti-rotatiestaaf draait u de anti-rotatiestaf evenwijdig ter hoogte van de zijkant, schuift u de abductiestaf over de beide anti-rotatiestaven en laat u deze op zijn plaats vasthouden. Als u de heupabductiestang wilt losmaken trekt u de heelendomhoog en schuift u de abductiestang van de anti-rotatiestaaf af.

AANBRINGEN EN VERWIJDEREN VAN DE LAARSZOO

Plaats voor het aanbrengen van de zoom de laarszoom op de onderkant van het frame, met de opening in de zoom ter hoogte van de draaiknop. Druk met de duimen aan weerszijden van de opening de zoom naar beneden en in de richting van de hiel. De draaiknop zal naar de achterkant van de zoomleugel schuiven. Druk de zoom bij de hand.

AFSTELLEN VAN HET TEENSTUK

Stel de lengte van het teenstuk in door de draaiknop aan de onderkant van de zoom los te maken. Plaats het teenstuk op de gevreesde lengte en draai de knop weer vast. Het teenstuk kan ook naar weerszijden van de voet worden gedraaid.

CONTRA-INDICATIES

Niet gebruiken bij patiënten met ernstige contractuur bij flexie van de voetzoom, een volledig of deelmatig gebrek aan de hiel.

WAARSCHUWING

• Inspectie of geen van de onderdelen van het hulpmiddel beschadigd is of ontbrekt, voordat het bij de patiënt wordt aangebracht.

• Raadpleeg onmiddellijk een arts of bevoegd medisch deskundige als u een veranderd gevoel ongewone reacties, zwelling of langdurig in een vinger of vinger tijdens het gebruik van dit product.

• Als het hulpmiddel niet wordt gebruikt in combinatie met een hulpmiddel voor sequentiële/intervallaire compressie, zorg er dan voor dat de slangen de ontlasting van de voet, enkel of kuit van de patiënt niet belemmeren.

• Gebruik het hulpmiddel niet en raadpleeg een arts of een veranderd gevoel als het hulpmiddel niet goed kan functioneren.

AANDACHTSPUNTEN

• Een arts of bevoegde beroepsbeoefenaar ("voorschrijver") die niet het gebruik en het doel dien van deze brace moet deze bij de gebruiker aanmetten. De voorschrijver is verplicht om voorzichtige enkele aanwijzingen te geven die de risico's aan te geven die verbonden zijn aan het gebruik van deze brace, evenals de duur van het gebruik. De instructies op dit blad niet bedoelen ter vervanging van het ziekenhuisprotocol of directe instructies van de voorschrijver.

• Controleer of de brace goed aansluit op de patiënt. **NIET TESTRAK**.

• Raadpleeg uw voorschrijver onmiddellijk als u een veranderd gevoel, ongewone reacties, zwelling of verhoogde pijn ervaart

BELANGRIKE INFORMATIE

• De klant kan alleen gebruik maken van de aanwijzingen en adviezen die door de leverancier van de hulpmiddel zijn gegeven.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel niet goed kan functioneren.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel niet goed kan functioneren.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel direct informeren als de hulpmiddel beschadigd is of ontbreekt.

• De klant moet de leverancier van de hulpmiddel