










# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling



symbol	explanatory text	standard reference	standard title
 <b>Catalogue or model number</b>	Indicates the manufacturer's catalogue number so that the medical device can be identified.	ISO 15223-1, Clause 5.1.6	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-2493	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Batch code</b>	Indicates the manufacturer's batch code so that the batch or lot can be identified.	ISO 15223-1, Clause 5.1.5	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-2492	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Serial number</b>	Indicates the manufacturer's serial number so that a specific medical device can be identified.	ISO 15223-1, Clause 5.1.7	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-2498	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Keep dry</b>	Indicates a medical device that needs to be protected from moisture.	ISO 15223-1, Clause 5.3.4	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-0626	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Keep away from sunlight</b>	Indicates a medical device that needs protection from light sources.	ISO 15223-1, Clause 53.2	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-0624	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Manufacturer</b>	Indicates the medical device manufacturer.	ISO 15223-1, Clause 5.1.1	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-3082	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Date of manufacture</b>	Indicates the date when the medical device was manufactured.	ISO 15223-1, Clause 5.1.3	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-2497	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Distributor</b>	Indicates the entity distributing the medical device into the locale.	ISO 15223-1, Clause 5.1.9	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-3724	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>EU Importer</b>	Indicates the entity importing the device into the locale.	ISO 15223-1, Clause 5.1.8	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
		ISO 7000-3725	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
 <b>Prescription only</b>	Requires prescription in the United States.	21 CFR 801.15(c)(1)(i)F	Labeling-Medical devices; prominence of required label statements.
		21 CFR 801.109	Labeling-Prescription devices.
 <b>Do not use if package is damaged</b>	Indicates a medical device that should not be used if the package has been damaged or opened.	ISO 15223-1, Clause 5.2.8	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2606	Graphical symbols for use on equipment.



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling



symbol	explanatory text	standard reference	standard title
	Indicates a medical device that can be broken or damaged if not handled carefully.	ISO 7000-0621, 2014-06-04	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
<b>Fragile, handle with care</b>			Graphical symbols for use on equipment.
	Indicates a medical device that contains a battery.	IEC 60417-5001B	Graphical symbols for use on equipment.
<b>Battery</b>			
	Indicates the presence of natural rubber or dry natural rubber latex as a material of construction within the medical device or the packaging of a medical device.	ISO 15223-1, Clause 5.4.5	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Contains natural rubber latex</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Indicates the materials of construction of the device or its packaging were not manufactured with natural rubber or dry natural rubber latex.	ISO 15223-1, Clause 5.4.5, Reference Annex B for negation symbol	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Not made with natural rubber latex</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied
	Indicates the date after which the medical device is not to be used.	ISO 15223-1, Clause 5.1.4	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Use by</b>		ISO 7000-2607	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Indicates the need for the user to consult the instructions for use.	ISO 15223-1, Clause 5.4.3	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Consult instructions for use</b>		IEC 60601-1, Table D.1, Symbol 11	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.
		ISO 7000-1641	Graphical symbols for use on equipment.
	Indicates the need for the user to consult the instructions for use for important cautionary information such as warnings and precautions that cannot, for a variety of reasons, be presented on the medical device itself.	ISO 15223-1, Clause 5.4.4	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Caution</b>		IEC 60601-1, Table D.1, Symbol 10	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.
	Indicates a medical device that is intended for one use or for use on a single patient during a single procedure. NOTE: Synonyms for “Do not re-use” are “single use” and “use only once”.	ISO 15223-1, Clause 5.4.2	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Do not reuse</b>		ISO 7000-1051	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Refer to instruction manual/ booklet.	IEC 60601-1, Table D.2, Symbol 10	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Follow instructions for use</b>			Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.
	Signifies European technical conformity.	765/2008/EC 768/2008/EC MDD 9 3/42/EEC Articles 4,11,12,17, Annex II )	The requirements for accreditation and market surveillance relating to the marketing of products; Medical Device Directive.
<b>CE marking</b>			
	Signifies UK (United Kingdom) technical conformity	Regulation 765/2008	The requirements for accreditation and market surveillance relating to marketing of products.
<b>UKCA Marking</b>			



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling



symbol	explanatory text	standard reference	standard title
	Indicates the authorized representative in the European Community.	ISO 15223-1, Clause 5.1.2	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Authorized European representative</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Indicates the temperature limits to which the medical device can be safely exposed.	ISO 15223-1, Clause 5.3.7	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Storage temperature range</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-0632	Graphical symbols for use on equipment.
	Indicates the lower limit of temperature to which the medical device can be safely exposed.	ISO 15223-1, Clause 5.3.5	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Lower limit of storage temperature</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-0534	Graphical symbols for use on equipment.
	Indicates the upper limit of temperature to which the medical device can be safely exposed.	ISO 15223-1, Clause 5.3.6	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Upper limit of storage temperature</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-0533	Graphical symbols for use on equipment.
	Indicates the range of humidity to which the medical device can be safely exposed.	ISO 15223-1, Clause 5.3.8	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Humidity Limitation</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2620	Graphical symbols for use on equipment.
	To indicate the acceptable upper and lower limits of atmospheric pressure for transport and storage.	ISO 15223-1, Clause 5.3.9	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Atmospheric Pressure</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2621	Graphical symbols for use on equipment.
	Indicates a medical device that is not to be re-sterilized.	ISO 15223-1, Clause 5.2.6	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Do not re-sterilize</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2608	Graphical symbols for use on equipment.
	Indicates a medical device that has not been subjected to a sterilization process.	ISO 15223-1, Clause 5.2.7	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Non sterile</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2609	Graphical symbols for use on equipment.
	Indicates a medical device that has been subjected to a sterilization process.	ISO 15223-1, Clause 5.2.1	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Product subjected to a sterilization process</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2499	Graphical symbols for use on equipment.
	Indicates a medical device that has been sterilized using irradiation.	ISO 15223-1, Clause 5.2.4	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Sterilized using irradiation</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2502	Graphical symbols for use on equipment.
	Indicates a medical device that has been sterilized using aseptic techniques.	ISO 15223-1, Clause 5.2.2	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Sterilized using aseptic techniques</b>			Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
		ISO 7000-2500	Graphical symbols for use on equipment.



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling



symbol	explanatory text	standard reference	standard title
	Indicates a medical device that has been sterilized using ethylene oxide.	ISO 15223-1, Clause 5.2.3	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Sterilized by ethylene oxide treatment</b>		ISO 7000-2501	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Indicates a medical device that has been sterilized using steam or dry heat.	ISO 15223-1, Clause 5.2.5	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Sterilized using steam or dry heat</b>		ISO 7000-2503	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Indicates there are potential biological risks associated with the medical device.	ISO 15223 5.4.1	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Biological Risks</b>		ISO 7000-0659	Medical devices — Symbols to be used with medical device labels, labeling and information to be supplied.
	Any terminal which is intended for connection to an external conductor for protection against electric shock in case of a fault, or the terminal of a protective earth (ground) electrode.	IEC 60417-5019	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Protected Earth; Protected Ground</b>			Graphical symbols for use on equipment. Used to mark the Protective Earthing Conductor terminal of Class I equipment.
	Double insulated appliance	IEC 60417-5172	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Double insulated appliance</b>			Graphical symbols for use on equipment.
	A Class III appliance is designed to be supplied from a separated extra-low voltage (SELV) power source. The voltage from a SELV supply is low enough that under normal conditions a person can safely come into contact with it without risk of electrical shock.	IEC 60417-5180	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>EE Class III</b>			Graphical symbols for use on equipment meeting the safety requirements specified for class III according to IEC 61140: Protection against electric shock — Common aspects for installation and equipment.
	Type B applied part complying with IEC 60601-1.	IEC 60417-5840	Medical Electrical Equipment - Part 1: General requirements for basic safety and essential performance.
<b>Type B applied part</b>		IEC 60878	
		IEC 60601-1	
	Type BF applied part	IEC 60417-5333	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Type BF applied part</b>			Graphical symbols for use on equipment.
	Classifies and rates the degree of protection provided by casing/ enclosure of electrical device.  <ul style="list-style-type: none"> <li>• <b>PROTECTION</b> against solid objects &gt; 12.5 mm;</li> <li>• <b>PROTECTION</b> against water sprayed up to 15 from vertical</li> </ul>	IEC 60529*	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>IP code, International Protection Marking</b>		*(does not have a reference number per 60601-1 [2005] p.645)	



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling












symbol	explanatory text	standard reference	standard title
	Magnetic resonance safe	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.
<b>MR Safe</b>			
	Magnetic resonance conditional	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.
<b>MR Conditional</b>			
	Magnetic resonance unsafe	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.
<b>MR Unsafe</b>			
	Defibrillation-Proof	IEC 60417-5336	Graphical symbols for use on equipment. Type CF of applied part.
<b>Defibrillation-proof</b>			
	Indicates part complies with higher degree of protection against electric shock as defined by IEC 60601-1.	IEC 60601-1 IEC 60878	Medical Electrical Equipment - Part 1: General requirements for basic safety and essential performance.
<b>EE Type CF</b>			
	To indicate that separate collection for waste electric and electronic equipment (WEEE) is required.	ISO 7000-6414 2012/19/EU	Graphical symbols for use on equipment. Waste Electrical and Electronic Equipment Directive
<b>WEEE</b>			
	Non-ionizing electromagnetic radiation	IEC 60417 REF 5140	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.
<b>Non-ionizing Radiation</b>			
	Indicates the materials of construction do not include the phthalate plasticizer DEHP	BS EN 15986 Annex B	Symbol for use in the labeling of medical devices – Requirements for labeling of medical devices containing phthalates.
<b>Does not contain DEHP</b>			
	Indicates the item is a medical device.	ISO 15223 – 1, Clause 5.7.7	Medical Devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General Requirements.
<b>Medical Device</b>			
	Indicates a single sterile barrier system.	ISO 7000-3707	Graphical symbols for use on equipment.
<b>Single Sterile Barrier System</b>			
	Indicates two sterile barrier systems.	ISO 7000-3704	Graphical symbols for use on equipment.
<b>Double Sterile Barrier System</b>			
	Indicates a single sterile barrier system with protective packaging inside.	ISO 7000-3708	Graphical symbols for use on equipment.
<b>Single Sterile Barrier System with Protective Packaging Inside</b>			



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling









symbol	explanatory text	standard reference	standard title
 <b>Single Sterile Barrier System with Protective Packaging Outside</b>	Indicates a single sterile barrier system with protective packaging outside.	ISO 7000-3709	Graphical symbols for use on equipment.
 <b>Benchmark Product Certification</b>	Benchmark mark is a registered certification mark owned and operated by BSI.	N/A	N/A
 <b>Accreditation Recognition Mark</b>	Compliance with AS/NZS (Australia/New Zealand) requirements.	N/A	N/A
 <b>NFPA 704 Hazard Diamond</b>	Provides a general idea of the hazards of a material and the severity of the hazards as they relate to emergency response. <ul style="list-style-type: none"> <li>• <b>RED:</b> Fire Hazard 0 - Will not burn</li> <li>• <b>BLUE:</b> Health Hazard 0 - Hazard no greater than ordinary material</li> <li>• <b>YELLOW:</b> Reactivity Hazard 0 - Stable</li> <li>• <b>WHITE:</b> Special - Eye Safety</li> </ul>	NFPA 704: 2022	National Fire Protection Association (NFPA): Standard System for the Identification of the Hazards of Materials for Emergency Response.
 <b>Alarm General</b>	To indicate an alarm on a control equipment	IEC 60601-1-8 and IEC 60417	Medical electrical equipment –  Part 1-8: General requirements for basic safety and essential performance—Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems and Graphical Symbols for Use on Equipment
 <b>Alarm Paused</b>	To identify the alarm inhibit on control equipment	IEC 60601-1-8 and IEC 60417	Medical electrical equipment –  Part 1-8: General requirements for basic safety and essential performance—Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems and Graphical Symbols for Use on Equipment
 <b>Audio Paused</b>	To identify the control for AUDIO PAUSED or to indicate that the ALARM SYSTEM is in the AUDIO PAUSED state	IEC 60601-1-8 and IEC 60417	Medical electrical equipment –  Part 1-8: General requirements for basic safety and essential performance—Collateral standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems and Graphical Symbols for Use on Equipment
 <b>Two Person Lift</b>	Indicates two person lift required	ANSI Z535.4 and ISO 3864-2	American National Standard for Product Safety Signs and Labels and Graphical symbols— Safety colours and safety signs, Part 2: Design principles for product safety labels
 <b>This Side Up</b>	To indicate correct upright position of the transport package	ISO 7000	Graphical symbols for use on equipment—Registered symbols



# Commonly Used Symbols

Glossary of symbols commonly used on  
DeRoyal Medical Device Packaging and Labeling



symbol	explanatory text	standard reference	standard title
 <b>Harmful</b>	This GHS pictogram with an exclamation mark indicates a substance that is harmful or irritating.	29 CFR 1910.1200	Occupational Safety and Health Standards—Hazard Communication
 <b>Warning</b>	General Warning Sign that requires a supplementary sign to give further information on the hazard. Risk specified by the supplementary sign.	ISO 7010	Graphical symbols—Safety colours and safety signs—Registered safety signs
 <b>Does not contain phthalates</b>	Indicates the materials of construction do not include phthalates	BS EN 15986 Annex B	Symbol for use in the labeling of medical devices—Requirements for labeling of medical devices containing phthalates
 <b>UL Classified Mark for Canada and US</b>	The UL Marks for Classification service appear on representative samples of products that UL Solutions has evaluated with respect to specific properties, a limited range of hazards or suitability for use under limited or special conditions.	N/A	N/A
 <b>Intertek ETL Listing Mark</b>	Intertek's ETL Mark is proof of product compliance to North American safety standards.	N/A	N/A
 <b>FCC Compliant</b>	This logo, or FCC mark, indicates that a device has been subject to testing and complies with FCC regulations.	47 CFR Part 18	Telecommunications—Federal Communications Commission—Industrial, Scientific, and Medical Equipment